

**4731.4426 SURVEYS OF AMBIENT RADIATION EXPOSURE RATE.**

A. In addition to the surveys required under parts 4731.2000 to 4731.2950, a licensee must survey with a radiation detection survey instrument at the end of each day of use all areas where unsealed radioactive materials requiring a written directive were prepared for use or administered.

B. A licensee need not perform the surveys required under item A in an area where patients or human research subjects are confined when they cannot be released under part 4731.4427.

C. A licensee must retain a record of each survey according to part 4731.4505.

**Statutory Authority:** *MS s 144.1202; 144.1203*

**History:** *29 SR 755*

**Published Electronically:** *March 12, 2009*