

**4731.4464 TREATMENT WITH REMOTE AFTERLOADER UNIT; SURVEYS.**

A. Before releasing a patient or human research subject who has been treated with a remote afterloader unit from licensee control, a licensee must survey the patient or human research subject and the remote afterloader unit with a portable radiation detection survey instrument to confirm that the source has been removed from the patient or human research subject and returned to the safe shielded position.

B. A licensee must retain a record of the required surveys according to part 4731.4511.

**Statutory Authority:** *MS s 144.1202; 144.1203*

**History:** *29 SR 755*

**Published Electronically:** *March 12, 2009*