

4731.4513 BRACHYTHERAPY SOURCE CALIBRATION RECORDS.

A licensee must maintain a record of the calibrations of brachytherapy sources required under part 4731.4455 for three years after the last use of the source. The records must include:

- A. the date of the calibration;
- B. the manufacturer's name, the model number and serial number for the source, and instruments used to calibrate the source;
- C. the source output or activity;
- D. the source positioning accuracy within the applicators; and
- E. the signature of the authorized medical physicist.

Statutory Authority: *MS s 144.1202; 144.1203*

History: *29 SR 755*

Published Electronically: *March 12, 2009*