

4731.4517 DOSIMETRY EQUIPMENT RECORDS.

A licensee must retain a record of the calibrations, intercomparisons, and comparisons of dosimetry equipment required under part 4731.4468 for the duration of the license. For each calibration, intercomparison, or comparison, the record must include:

- A. the date;
- B. the manufacturer's name, model number, and serial number for the instrument that was calibrated, intercompared, or compared;
- C. the correction factor that was determined from the calibration or comparison or the apparent correction factor that was determined from an intercomparison; and
- D. the name of the individual who performed the calibration, intercomparison, or comparison.

Statutory Authority: *MS s 144.1202; 144.1203*

History: *29 SR 755*

Published Electronically: *March 12, 2009*