

**4731.4518 CALIBRATION RECORDS; TELETHERAPY, REMOTE AFTERLOADER, AND GAMMA STEREOTACTIC RADIOSURGERY UNITS.**

A licensee must maintain a record of the teletherapy unit, remote afterloader unit, and gamma stereotactic radiosurgery unit full calibrations required under parts 4731.4469 to 4731.4471 for three years. The record must include:

- A. the date of the calibration;
- B. the unit manufacturer's name;
- C. the model number and serial number of the teletherapy, remote afterloader, and gamma stereotactic radiosurgery unit;
- D. the model number, serial number, and identity of the source;
- E. the model number, serial number, and identity of the source instruments used to calibrate the units;
- F. the results and an assessment of the full calibrations;
- G. the results of the autoradiograph required for low dose-rate remote afterloader units; and
- H. the signature of the authorized medical physicist who performed the full calibration.

**Statutory Authority:** *MS s 144.1202; 144.1203*

**History:** *29 SR 755*

**Published Electronically:** *March 12, 2009*