

4731.4520 SPOT CHECK RECORDS; REMOTE AFTERLOADER UNITS.

A. A licensee must retain a record of each spot check for remote afterloader units required under part 4731.4473 for three years. The record must include, as applicable:

- (1) the date of the spot check;
- (2) the unit manufacturer's name;
- (3) the model number and serial number for the remote afterloader unit;
- (4) the model number, serial number, and identity of the source;
- (5) an assessment of timer accuracy;
- (6) notations indicating the operability of each entrance door electrical interlock, radiation monitor, source exposure indicator light, viewing and intercom system, and clock and decayed source activity in the unit's computer;
- (7) the name of the individual who performed the periodic spot check; and
- (8) the signature of the authorized medical physicist who reviewed the record of the spot check.

B. A licensee must retain a copy of the procedures required under part 4731.4473, subpart 2, until the licensee no longer possesses the remote afterloader unit.

Statutory Authority: *MS s 144.1202; 144.1203*

History: *29 SR 755*

Published Electronically: *March 12, 2009*