

4731.4522 OPERABILITY RECORDS; MOBILE REMOTE AFTERLOADER UNITS.

A licensee must retain a record of each check for mobile remote afterloader units required under part 4731.4475 for three years. The record must include:

- A. the date of the check;
- B. the manufacturer's name, model number, and serial number for the remote afterloader unit;
- C. notations accounting for all sources before the licensee departs from a facility;
- D. notations indicating the operability of each entrance door electrical interlock, radiation monitor, source exposure indicator light, viewing and intercom system, applicator, source transfer tube, and transfer tube applicator interface and the source positioning accuracy; and
- E. the signature of the individual who performed the check.

Statutory Authority: *MS s 144.1202; 144.1203*

History: *29 SR 755*

Published Electronically: *March 12, 2009*