

4731.4524 FULL-INSPECTION SERVICING RECORDS; TELETHERAPY AND GAMMA STEREOTACTIC RADIOSURGERY UNITS.

A licensee must maintain a record of the full-inspection servicing for teletherapy and gamma stereotactic radiosurgery units required under part 4731.4477 for the duration of use of the unit. The record must contain:

- A. the inspector's radioactive material license number;
- B. the date of inspection;
- C. the manufacturer's name, model number, and serial number for both the treatment unit and source;
- D. a list of components inspected and serviced and the type of service; and
- E. the signature of the inspector.

Statutory Authority: *MS s 144.1202; 144.1203*

History: *29 SR 755; 46 SR 791*

Published Electronically: *May 26, 2022*