

**4732.0710 RADIATION SURVEY OR MEASUREMENT INSTRUMENTS.**

Subpart 1. **Requirements.** To ensure correct response to radiation, each radiation survey instrument must be calibrated at intervals not to exceed 24 months and after each servicing:

- A. be calibrated at energy levels and over a range appropriate for the use;
- B. be calibrated to accuracy within plus or minus 20 percent over the applicable range of the instrument;
- C. have records of the calibrations maintained according to part 4732.0330;
- D. the calibration of any electronic equipment must be traceable to its calibration standard at the National Institute of Standards and Technology (NIST); and
- E. noninvasive kVp meters must be calibrated by the manufacturer or an accredited calibration laboratory.

Subp. 2. **Records.** The registrant must maintain the records of the tests and calibrations according to part 4732.0330.

**Statutory Authority:** *MS s 144.12*

**History:** *32 SR 777*

**Published Electronically:** *December 10, 2007*